



Sound Level Meter Calibration Report

Sound Level Meter Manufacturer:	Galaxy	Model Number:	CM-140
Serial Number:	100907672	Frequency Weighting:	A & C-Weighting
Measurement Date:	March 26, 2010	Temperature:	66°F/19°C
Engineer:	H. Singleton	Humidity:	28%
Ref. Calibrator:	GenRad 1986		
Serial Number:	00244		
Calibration Level (C-Weighted):	94.0 dB (0 dB deviation)		
(re 1000 Hz, 94 dB)			
Calibration Level (A-Weighted):	94.0 dB (0 dB deviation)		
(re 1000 Hz, 94 dB)			